

PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

K-BAR 2000B Measuring System

manufactured by:

Kurz Instruments Inc.

2411 Garden Road Monterey CA 93940 USA

has been assessed by Sira Certification Service and for the conditions stated on this certificate complies with:

Environment Agency Guidance "MCERTS for stack emissions monitoring equipment at industrial installations" - Continuous emissions monitoring systems(CEMS) Published 20 October 2020 EN15267-1, EN15267-2, EN15267-3, EN ISO 16911-2 & QAL 1 as defined in EN 14181: 2014

Certification ranges:

Velocity (

0 to 30m/s

Project number: Certificate number: Initial certification: This certificate issued: Renewal date: 800059169 Sira MC150275/01 17 June 2015 14 December 2020 16 June 2025

Andrew Young Environmental Team Manager

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service



Unit 6, Hawarden Industrial Park Hawarden, Deeside, CH5 3US Tel: +44 (0)1244 670 900



The MCERTS certificate consists of this document in its entirety. For conditions of use, please consider all the information within. This certificate may only be reproduced in its entirety and without change To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts







Certificate contents

Approved site application	2
Basis of certification	2
Product certified	
Certified performance	
Description	
General notes	

Approved site application

Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency technical guidance on monitoring, available at <u>www.mcerts.net</u>

This instrument is considered suitable for use on waste incineration and large combustion plant applications. This CEMS has been proven suitable for its measuring task (parameter and composition of the flue gas) by use of the QAL 1 procedure specified in EN14181. The lowest certified range for each determinand shall not be more than 1.5 times the daily average emission limit value (ELV) for incineration plants, and not more than 2.5 times the ELV for other types of application.

Field test was conducted on a municipal waste incinerator for 3 months.

Basis of certification

This certification is based on the following test report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

TÜV report number 936/2121960/A dated 10th October 2013

Product certified

The K-BAR 2000B measuring system consists of the following parts:

- K-BAR 2000B multipoint insertion flow element(s) for duct under measurement.
- Adam 155 flow computer to support all flow elements of the K-BAR.
- Kurz AMS-PLC to provide data quality status.

This certificate applies to all instruments fitted with software version MFT-B v2.08 or higher (Kurz serial number 1294A onwards).

Certificate number: Sira MC150275/01 This certificate issued: 14 December 2020







Certified performance

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range:	-20°C to +50°C
Instrument IP rating:	IP66

Note: If the instrument is supplied with an enclosure, then the ambient temperature shall be monitored inside the enclosure to ensure that it stays within the above ambient temperature range.

Results are expressed	as error % certification range,	unless otherwise stated.

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Response time						
Velocity					2s	<60s
Repeatability standard deviation at zero point						
Velocity	0.1					<2.0%
Repeatability standard deviation at reference point						
Velocity	0.3					<2.0%
Lack-of-fit						
Velocity			-1.33			<3.0%
Influence of ambient temperature zero point						
Velocity	-0.2					<5.0%
Influence of ambient temperature reference point						
Velocity		-0.7				<5.0%
Influence of voltage variations 190 to 250V						
Velocity	0.1					<2.0%
Measurement uncertainty						
Velocity					2.7	<7.5% (10%)
Calibration function (field)					Note 1	
Velocity						>0.90
Response time (field)						
Velocity					2s	<200s

Certificate number: This certificate issued: Sira MC150275/01 14 December 2020

This certificate may only be reproduced in its entirety and without change To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts







Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Maintenance interval					Note 2	
					4 Weeks	>8 days
Zero and Span drift requirement	A voltage is substituted at a zero, mid and span values to validate the signal from the sensor control board in the K-BAR, thru the 155 flow computer to the 4-20 mA output monitored by the customer. This is an "Electronic Zero-Mid-Span" test. Actual flow testing must be done with a standard traverse method or relative accuracy test on a periodic basis to confirm the sensor stability and suitability of the periodic maintenance cleaning of the sensor. Deviations from the proper reading are corrected by repair or change to the meter's correction factor in the 155 flow computer.					Clause 6.13 & 10.13 Manufacturer shall provide a description of the technique to determine and compensate for zero and span drift.
Change in zero point over maintenance interval						
Velocity	0.2					<2.0%
Change in reference point over maintenance interval						
Velocity	0.3					<4.0%
Availability						
					99.9%	>95%
Reproducibility						
Velocity			1.4			<3.3%

Note 1: Variation of readings <15% of certification range. Therefore R2 calculation is not necessary. The instrument passed variability testing.

Note 2: The K-BAR 2000B measuring system has a maintenance interval of 4 weeks. The work detailed below has to be carried out at regular intervals, depending on local conditions:

- Regular visual inspection.
- Reference point control using the control cycle.
- Ensure manufacturer instructions are always followed.

Certificate number: Sira MC150275/01 This certificate issued: 14 December 2020

> This certificate may only be reproduced in its entirety and without change To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts







Description

K-BAR 2000B measuring system uses up to four sensors for measurement redundancy and to ensure accuracy. The instrument is designed to withstand the stress and vibrations found in large industrial stacks and ducts that commonly have wide-ranging velocity and temperature profiles. The K-BAR 2000B is designed for process temperatures from the dew point up to 260°C or up to 500°C.

General notes

- 1. This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this certificate. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations applicable to the holders of Sira certificates'.
- 2. The design of the product certified is held and maintained by TUV Rheinland for certificate No. Sira MC150275/01.
- 3. If a certified product is found not to comply, Sira should be notified immediately at the address shown on this certificate.
- 4. The certification marks that can be applied to the product or used in publicity material are defined in 'Regulations applicable to the holders of Sira certificates'.
- 5. This document remains the property of Sira and shall be returned if requested by Sira.