

Certificate



No.: 968/FSP 1872.01/24

Product tested	B-Series Flow Meter to measure the mass flow rate of gases	Certificate holder	Kurz Instruments Inc. 2411 Garden Road 93940 Monterey USA
Type designation	454FTB, 454FTB-WGF, 534FTB, 504FTB, K-BAR 2000B and K-BAR 2000B-WGF		
Codes and standards	IEC 61508 Parts 1-7:2010		
Intended application	The products comply with the requirements of SIL 1 acc. to IEC 61508 and can be used in safety related applications up to these safety levels.		
Specific requirements	The instructions of the associated Installation, Operating and Safety Manual shall be considered.		
Valid until 2029-06-05			

The issue of this certificate is based upon an examination, whose results are documented in Report No. 968/FSP 1872.01/24 dated 2024-05-16.
This certificate is valid only for products which are identical with the product tested.

TÜV Rheinland Industrie Service GmbH
Bereich Automation
Funktionale Sicherheit
Am Grauen Stein, 51105 Köln

Köln, 2024-06-05

Certification Body Safety & Security for Automation & Grid

Dipl.-Ing. Thomas Steffens